
BHASKARACHARYA COLLEGE OF APPLIED SCIENCES

(University of Delhi)



E-Procurement Tender Notice

The College invites ONLINE bids as per Two bids System (Technical and Financial) from eligible bidders through e-procurement https://eprocure.gov.in/eprocure/app

Our Enquiry Ref No : BCAS/Bot/NR/Quot./2018/						
Dated	:	_28/02/2018_				
Tender fee (Rs.)	:	500/-				
Bid download start Date and Time	:	28/02/2018, -1530 hrs				
Bid submission start Date and Time	:	28/02/2018,- 1600 hrs				
Last date and time for Bid Submission	:	21/03/2018-, 1600 hrs				
Date and Time of Bid Opening	:	22/03/2018-, 1600 hrs				
Tender Value (Rs.)	:	3,65,000/-				
EMD(Rs.)	:	7,500/-				
Bid Validity	:	Up to 31.03.2018				

Subject: Invitation of ONLINE 2 fold bids for the Procurement of Non-Recurring Items by the Department of Botany.

Dear Bidder,

Only the online bids are invited for the procurement of Laboratory Equipment / Spectrophotometer as per the details attached as Enclosure-I of the item(s).

S. No.		Max. Required	Quantity	Minimum Specifications
1	Rotary Microtome Automatic	One		Enclosure -I

IMPORTANT:

- A11 details regarding the subject tender are available on websites www.bcas.du.ac.in and https://eprocure.gov.in/eprocure/app. Any change/ modification in the Tender Enquiry/ Tender Document will be intimated through above websites only. Bidders are therefore, requested to visit the websites regularly to keep themselves updated.
- Bids shall be submitted online only at CPPP website: http://eprocure.gov.in/eprocure/app
- Manual bids shall not be accepted.
- For submission of E-Bids, bidders registered with are required get themselves to http://eprocure.gov.in/eprocure/app
- Bidder advised to follow the instructions provided in the 'Instructions to the Contractors/Bidder' for the esubmission of the bids online through the Central Public Procurement Portal for e Procurement at https://eprocure.gov.in/eprocure/app
- Bid documents may be scanned with minimum 100dpi with black and white option in pdf format.

It is required that the following instructions should be carefully followed including detailed terms and conditions attached overleaf as Annexure 'A', while submitting your offer; otherwise your offer may not be considered.

- All the communication with the college should be addressed only to "Principal, Bhaskaracharya College of 1. Applied Sciences, Sector 2, Phase I, Dwarka, New Delhi- 110 075." (hereinafter called the Principal)
- 2. Online Quotations will be two fold (a) one technical bid consisting of all technical details and supporting documents (b) another financial bid containing items wise price for the items mentioned in the technical bid. Bidders will not be permitted to alter or modify their bids after expiry of the deadline for receipt of bids.
- Financial bids of only those bidders will be opened and considered who qualify in their technical bid. 3.
- 4. Corrigendum, if any, will be published only on the above websites only.

Yours Sincerely,

Principal

Enclosure – I

S.No	Name of Item	Technical Specification	Quantity Required
1	Rotary Microtome - Automatic	Range of section thickness: 0.5 - 100 µm from 0.5 µm - 2 µm in 0.5 µm increment ,from 2 µm - 10 µm in 1 µm increment, from 10µm- 20µm in 2µm increment & 20µm -100µm in 5µm increment Section Thickness Indication: Bright LCD Trimming Thickness: 1 - 600 µm, from 1 µm - 10 µm in 1 µm increment, from 10µm- 20µm in 2µm increment & 20µm - 100µm in 10µm increment & from 100µm - 600µ in 50µm increment Specimen Retraction: 20µm Object feed : 28mm Vertical stroke: 60mm Specimen Clamp Orientation : 360° Maximum Specimen Size: 50mmX50mm Operating Voltage : 220V Single Phase AC supply	ONE